

***TB 9-6625-2263-35**

DEPARTMENT OF THE ARMY TECHNICAL BULLETIN

CALIBRATION PROCEDURE FOR DIGITAL MULTIMETER JOHN FLUKE, MODEL 8520A

Headquarters, Department of the Army, Washington, DC

13 August 2003

Approved for public release; distribution is unlimited

REPORTING OF ERRORS AND RECOMMENDING IMPROVEMENTS

You can improve this manual. If you find any mistakes or if you know of a way to improve these procedures, please let us know. Mail your letter or DA Form 2028 (Recommended Changes to Publications and Blank Forms) directly to: Commander, U.S. Army Aviation and Missile Command, ATTN: AMSAM-MMC-MA-NP, Redstone Arsenal, AL 35898-5000. A reply will be furnished to you. You may also provide DA Form 2028 information to AMCOM via email, fax, or the World Wide Web. Our fax number is DSN 788-6546 or Commercial 256-842-6546. Our e-mail address is: 2028@redstone.army.mil. Instructions for sending an electronic 2028 may be found at the back of this manual. For the World Wide Web, use <https://amcom2028.redstone.army.mil>.

		Paragraph	Page
SECTION	I. IDENTIFICATION AND DESCRIPTION		
	Test instrument identification	1	2
	Forms, records, and reports	2	2
	Calibration description	3	2
	II. EQUIPMENT REQUIREMENTS		
	Equipment required.....	4	3
	Accessories required.....	5	3
	III. CALIBRATION PROCESS		
	Preliminary instructions.....	6	5
	Equipment Setup.....	7	5
	Dc voltage	8	6
	Resistance.....	9	9
	Ac voltage.....	10	11
	Power supply.....	11	12
	Final procedure	12	13

*This bulletin supersedes TB 9-6625-2263-35 dated 14 January 1992.

SECTION I IDENTIFICATION AND DESCRIPTION

1. Test Instrument Identification. This bulletin provides instructions for the calibration of Digital Multimeter, John Fluke, Model 8520A. The manufacturer's manual was used as the prime data source in compiling these instructions. The equipment being calibrated will be referred to as the TI (test instrument) throughout this bulletin.

a. Model Variations. None.

b. Time and Technique. The time required for this calibration is approximately 2 hours, using the dc and low frequency technique.

2. Forms, Records, and Reports

a. Forms, records, and reports required for calibration personnel at all levels are prescribed by TB 750-25.

b. Adjustments to be reported are designated (R) at the end of the sentence in which they appear. When adjustments are in tables, the (R) follows the designated adjustment. Report only those adjustments made and designated with (R).

3. Calibration Description. TI parameters and performance specifications which pertain to this calibration are listed in table 1.

Table 1. Calibration Description

Test Instrument Parameters	Performance Specifications		
Dc voltage			Accuracy ±(% of input + digits)
	Range	Full-scale	
	100 mV	199.999	0.011 + 10
	1 V	1.99999	0.011 + 2
	10 V	16.0100	0.009 + 1
	100 V	130.100	0.012 + 2
	1000 V	1024.00	0.011 + 1
Ac voltage			
	Range	Full-scale	Accuracy ¹ ±(% of input + % of FS)
	1 V	1.99999	10 to 20 Hz 3.5 + 0.6
	10 V	16.0100	20 to 40 Hz 0.6 + 0.6
	100 V	130.100	40 Hz to 20 kHz 0.15 + 0.05
	650 V	650.00	20 to 100 kHz 2.0 + 0.6
			100 to 300 kHz 4.0 + 1.0
			300 kHz to 1 MHz 15.0 + 5.0

See footnotes at end of table.