

TECHNICAL MANUAL

OPERATOR'S, ORGANIZATIONAL,

DIRECT SUPPORT, AND GENERAL SUPPORT

MAINTENANCE MANUAL

INCLUDING REPAIR PARTS AND

SPECIAL TOOLS LISTS

VOLTMETER, DIGITAL AN/GSM-64B

(NSN 6625-00-022-7894)

This copy is a reprint which includes current
pages from Change 1.

HEADQUARTERS, DEPARTMENT OF THE ARMY

MAY 1975

TECHNICAL MANUAL
No. 11-6625-444-14-1

HEADQUARTERS
DEPARTMENT OF THE ARMY
WASHINGTON, DC., 1 May 1975.

**Operator's, Organizational, Direct Support,
and General Support Maintenance Manual
Including Repair Parts and Special Tools Lists**

VOLTMETER, DIGITAL AN/GSM-64B

(NSN 6625-00-022-7894)

INCLUDING

PLUG-IN, ELECTRONIC TEST EQUIPMENT

PL-1370/GSM-64B

(NSN 6625-00-137-8366)

Current as of February 1975.

		Paragraph	Page
CHAPTER	1. INTRODUCTION		
Section	I. General	1-1-1-4.3	1-1
	II. Description and data	1-5-1-8	1-1
CHAPTER	2. INSTALLATION		
Section	I. Service upon receipt of material	2-1,2.2	2-1
	II. Installation instructions	2-3,2-4	2-2
CHAPTER	3. OPERATING INSTRUCTIONS		
Section	I. Controls and instruments	3-1-3.4	3-1
	II. Operation under unusual conditions	3-5-3-12	3-5
	III. Preparation for movement	3-13-3-15	3-7
CHAPTER	4. OPERATOR'S AND ORGANIZATIONAL, MAINTENANCE		
Section	I. Tools and equipment	4-1,4-2	4-1
	II. Preventive maintenance checks and services	4-3,4-4	4-1
	III. Troubleshooting	4-5,4-6	4-2
	IV. Maintenance of voltmeter	4-7-4-10	4-3
CHAPTER	5. FUNCTIONING OF EQUIPMENT		
Section	I. Block diagram analysis	5-1-5-3	5-1
	II. Circuit description	5-4-5-13	5-4
CHAPTER	6. GENERAL SUPPORT MAINTENANCE		
Section	I. General	6-1-6-4	6-1
	II. Tools, test equipment and troubleshooting	6-5-6-8	6-15
	III. Maintenance of the voltmeter	6-9-6-14	6-23
	IV. Testing procedures	6-15-6-17	6-27
CHAPTER	7. MATERIEL USED IN CONJUNCTION WITH AN/GSM-64B		
Section	I. Introduction to PL-1370/GSM-64B	7-1-7-4	7-1
	II. Operation of PL-1370/GSM-64B	7-5-7-9	7-2
	III. Functioning of PL-1370/GSM-64B	7-10-7-12	7-3
	IV. Troubleshooting and repair of PL-1370/GSM-64B	7-13-7-18	7-6

	Paragraph	Page
APPENDIX A	REFERENCES	A-1
	B. OPERATOR'S ORGANIZATIONAL, DIRECT SUPPORT AND GENERAL SUPPORT MAINTENANCE REPAIR PARTS ANDS SPECIAL TOOLS LISTS	
SECTION I.	Introduction	B-1
II	Basic issue items list	B-3 1
III.	Items troop installed or authorized list	B-4
IV.	Repair parts list	B-5
GROUP 0101	Digital Voltmeter AN/GSM-64B	B-5 2
0102	Power supply assembly	B-5 3
0103	Range supply assembly	B-7 4
0104	Ratio input assembly	B-8 5
0105	Logic, PCB	B-9 6
0106	A-D converter assembly	B-10 7
0107	Active filter assembly	B-12 8
0108	Buffer assembly	B-13 9
0109	Final assembly	B-14 10
0110	Front panel assembly	B-15 11
0111	Rear panel assembly	B-16 12
0112	Display assembly	B-16 13
0113	Decimal logic assembly	B-18 14
0114	Plug-in Electronic Test Equipment PL-1370/GSM-64B	B-18.1 15
Section V.	Special tools list (Not applicable)	
VI.	National stock number and part number index	B-19
APPENDIX	C. MAINTENANCE: ALLOCATION	
Section I.	Introduction	C-1
II.	Maintenance allocation chart	C-3

INDEX	Index 1
-------	---------

LIST OF ILLUSTRATIONS


Figure number	Title	Page
1 - 1	Digital voltmeter AN/GSM-64B	1-0
1 - 2	Front panel	1-3
1 - 3	Rear panel	1-3
1 - 4	Outline drawing	1-3
2 - 1	Packaging diagram	2-2
2 - 2	Rack installation	2-3
2 - 3	Equipment connections	2-3
3 - 1	Guard connections	3-2
3 - 2 ①	Controls, connectors, and indicators (front panel)	3-4
3 - 2 ②	Controls, connectors, and indicators (rear panel)	3-4
3 - 3	Operational check equipment connections	3-5
3 - 4	Ratio overrange limits	3-7
5 - 1	Voltmeter block diagram	5-3
5 - 2	Buffer block diagram	5-4
5 - 3	Active filter block diagram	5-5
5 - 4	A/D converter block diagram	5-7
5 - 5 ①	Timing and synchrony diagram (sheet 1 of 3)	5-10
5 - 5 ②	Timing and synchrony diagram (sheet 2 of 3)	5-11
5 - 5 ③	Timing and synchrony diagram (sheet 3 of 3)	5-12
5 - 6	Basic sample command oscillator	5-13
5 - 7	Decimal logic truth tables	5-17
5 - 8	Decimal logic (A18), schematic diagram	5-18
6 - 1	Assembly and troubleshooting test point location	6-4
6 - 2	Alignment adjustment and test point location	6-5
6 - 3	Buffer (A11) component parts	6-6
6 - 4	Active filter (A10) component parts	6-7
6 - 5	A-to-D converter (A9) component parts	6-8
6 - 6	Logic (A8) component parts	6-9
6 - 7	Display (A14) component parts	6-10
6 - 8	Range delay (A2) component parts	6-11
6 - 9	Power supply (A1) component parts	6-12

<i>Figure number</i>	<i>Title</i>	<i>Page</i>
6-10	Ratio input (A3) component parts	6-13
6-11	Decimal logic (A18) component parts	6-14
6-12	Front panel (A16) component parts	6-14
6-13	Rear panel (A17) component parts	6-15
6-14	A-to-D converter and logic, troubleshooting	6-17
6-15	Display troubleshooting	6-18
6-16	A-to-D converter waveforms	6-19
6-17	Analog storage waveforms	6-19
6-18	16-State binary counter waveforms	6-19
6-19	6-State shift register waveforms	6-20
6-20	A9TP2 zero input	6-20
6-21	A9TP2 6.3524V input	6-21
6-22	A9TP2 0.99999V input	6-21
6-23	Disassembly sequence (a through m)	6-24
6-24	Disassembly sequence (n through u)	6-25
6-25	Ratio check equipment connections	6-28
6-26	Alignment equipment connections	6-29
7-1	Ac converter, location diagram	7-2
7-2	Ac converter, block diagram	7-4
7-3	Plug-in, Electronic Test Equipment PL- 1370 /GSM-64B	7-7
6-26	Alignment equipment connections	6-29
1	Extender card	B-27
2	Digital voltmeter AN/GSM-64B	B-28
3	Power assembly	B-29
4	Range delay assembly	B-30
5	Ratio input assembly	B-31
6	Logic assembly	B-32
7	A-to-D converter assembly	B-33
8	Active filter assembly	B-34
9	Buffer amplifier assembly	B-35
10	Chassis assembly	B-36
11	Front panel assembly	B-37
12	Rear panel assembly	B-38
13	Display assembly	B-39
14	Decimal logic assembly	B-40
FO-1	Color code marking for MIL STD resistors, inductors, and capacitors	Fold-in
FO-2	Buffer (A11) schematic diagram	Fold-in
FO-3	Active filter (A10) schematic diagram	Fold-in
FO-4	A-to-D converter (A9) schematic diagram	Fold-in
FO-5	Logic (A8) schematic diagram	Fold-in
FO-6	Display (A14) schematic diagram (sheet 1 of 2)	Fold-in
FO-6	Display (A14) schematic diagram (sheet 2 of 2)	Fold-in
FO-7	Range delay (A2) schematic diagram	Fold-in
FO-8	Power supply (A 1) schematic diagram	Fold-in
FO-9	Ratio input (A3) schematic diagram	Fold-in
FO-10	Interconnect diagram	Fold-in
FO-11	Wiring diagram	Fold-in
FO-12	AC converter, schematic diagram	Fold-in

LIST OF TABLES

<i>Number</i>	<i>Title</i>	<i>Page</i>
3-1	Controls and indicators	3-2
3-2	DC ratio measurements	3-6
3-3	Data drive signals	3-6
4-1	Materials required	4-1
4-2	Operator's preventive maintenance checks and services	4-2
4-3	Organizational preventive maintenance checks and services	4-2
5-1	Sample and hold switching	5-8
5-2	Shift register outputs	5-9
5-3	Sixteen-state binary output	5-9
6-1	Power supply (A1) voltage measurements	6-1
6-2	Range delay (A2) voltage measurements	6-1

Number	Title	Page
6-3	Ratio input (A3) voltage measurements	6-2
6-4	Logic (A8) voltage measurements	6-2
6-5	Active filter (A10) voltage measurements	6-3
6-6	Buffer (A11) voltage measurements	6-3
6-7	Required test equipment	6-15
6-8	Power supply (A1) checks	6-16
6-9	Digital section subsections	6-16
6-10	Buffer (A11) resistor values	6-27
6-11	Voltmeter accuracy checks	6-28
6-12	Ladder alignment	6-30
6-13	Comparator level adjustment	6-30
6-14	Buffer DC alignment	6-31
7-1	Performance checks	7-6
7-2	Voltage measurements	7-8
7-3	Ac converter function command check	7-8
7-4	Ac converter relay truth table	7-8



EL6625-444-14-TM-1

Figure 1-1. Digital Voltmeter AN/GSM-64B.

CHAPTER 1

INTRODUCTION

Section I. GENERAL

1-1. Scope

This manual describes Voltmeter, Digital AN/GSM-64B (voltmeter) (fig. 1-1) and provides instructions for operation and organizational and general support maintenance. Direct support maintenance is not authorized for the AN/GSM-64B. Instructions are provided for the operator and the organizational repairman for installation, operation, preventive maintenance, and replacement of parts available at organizational maintenance. Circuit functioning is included for general support maintenance, together with instructions appropriate to this category of maintenance for troubleshooting, testing, adjusting, aligning and repairing the equipment and replacement of maintenance parts. Chapter 7 describes optional plug-in subassemblies which will extend the functional capability of the voltmeter. The options are not general issue, but are available to those maintenance facilities authorized their use.

1-2. Indexes of Publications

a. *DA Pamphlet 310-4*. Refer to the latest issue of DA Pam 310-4 to determine whether there are new editions, changes, or additional publications pertaining to the equipment.

b. *DA Pamphlet 310-7*. Refer to DA Pam 310-7 to determine whether there are modification work orders (MWO's) pertaining to the equipment.

1-3. Forms and Records

a. *Reports of Maintenance and Unsatisfactory Equipment*. Maintenance forms, records, and reports which are to be used by maintenance personnel at all maintenance levels are listed in and prescribed by TM 38-750.

b. *Report of Packaging and Handling Deficiencies*. Fill out and forward DD Form 6 (packaging Improvement Report) as prescribed in

AR 700-58/NAV SUPINST 4030.29/AFR 71-13/MCO P4030.29A, and DSAR 4145.8.

c. *Discrepancy in Shipment Report (DISREP) (SF 361)*. Fill out and forward Discrepancy in Shipment Report (DISREP) (SF 361) as prescribed in AR 55-38/NAV SUPINST 4610.36A/AFR 75-18/MCO P4610.19B, and DSAR 4500.15.

1-4. Recommendation for Maintenance Publication Improvements

Report of errors, omissions, and recommendations for improving this publication by the individual user is encouraged. Reports should be submitted on DA Form 2028 (Recommended Changes to Publications and Blank Forms) and forwarded direct to Commander, US Army Electronics Command, ATTN: DRSEL-MA-Q, Fort Monmouth, NJ 07703.

1-4.1. Administrative Storage

For procedures, forms, and records, and inspections required during administrative storage of the equipment, refer to TM 740-90-1.

1-4.2. Destruction of Army Materiel

Demolition and destruction of electronic equipment will be under the direction of the Commander and in accordance with TM 750-244-2.

1-4.3. Reporting Equipment Improvement Recommendations (EIR).

EIR will be prepared using DA Form 2407, Maintenance Request. Instructions for preparing EIR's are provided in TM 38-750, The Army Maintenance Management System. EIR's should be mailed directly to Commander, US Army Electronics Command, ATTN: DRSEL-MA-Q, Fort Monmouth, New Jersey 07703. A reply will be furnished directly to you.

Section II. DESCRIPTION AND DATA

1-5. Purpose and use

a. The voltmeter is a precision digital instrument capable of accurately measuring voltages

from 0 to 1,100 vdc. It also allows voltage ratio comparison between the voltage input and an external reference voltage from ± 11.0 to 100 vdc.